



Measurement of oxygen transfer from air into organic solvents

Oxygen transfer from air into organic solvents

Ramesh, Hemalata; Mayr, Torsten; Hobisch, Mathias; Borisov, Sergey; Klimant, Ingo; Krühne, Ulrich; Woodley, John M

Published in:
Journal of Chemical Technology and Biotechnology

Link to article, DOI:
[10.1002/jctb.4862](https://doi.org/10.1002/jctb.4862)

Publication date:
2016

Document Version
Peer reviewed version

[Link back to DTU Orbit](#)

Citation (APA):
Ramesh, H., Mayr, T., Hobisch, M., Borisov, S., Klimant, I., Krühne, U., & Woodley, J. M. (2016). Measurement of oxygen transfer from air into organic solvents: Oxygen transfer from air into organic solvents. *Journal of Chemical Technology and Biotechnology*, 91(3), 832–836. <https://doi.org/10.1002/jctb.4862>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Measurement of oxygen transfer from air into organic solvents

Hemalata Ramesh¹, Torsten Mayr^{2,*}, Mathias Hobisch², Sergey Borisov², Ingo Klimant², Ulrich Krühne¹ and John M Woodley¹

¹Department of Chemical and Biochemical Engineering, Technical University of Denmark, DK-2800 Kgs. Lyngby, Denmark

²Institute of Analytical Chemistry and Food Chemistry, Graz University of Technology, 8010 Graz, Austria

Correspondence: Assoc. Prof. Torsten Mayr, Institute of Analytical Chemistry and Food Chemistry, Graz University of Technology, 8010 Graz, Austria.

E-mail: torsten.mayr@tugraz.at

This article has been accepted for publication and undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the Version of Record. Please cite this article as doi: 10.1002/jctb.4862

Abstract**Background**

The use of non-aqueous organic media is becoming increasingly important in many biotechnological applications in order to achieve process intensification. Such media can be used for example to directly extract poorly water-soluble toxic products from fermentations. Likewise many biological reactions require the supply of oxygen, most normally from air. However, reliable on-line measurements of oxygen concentration in organic solvents (and hence oxygen transfer rates from air to the solvent) has to date proven impossible due limitations in the current analytical methods.

Results

For the first time, we demonstrate on-line oxygen measurements in non-aqueous media using a novel optical sensor. The sensor was used to measure oxygen concentration in various organic solvents including toluene, THF, isooctane, DMF, heptane and hexane (which have all been shown suitable for several biological applications). Subsequently, we measured the oxygen transfer rates from air into these organic solvents.

Conclusion

The measurement of oxygen transfer rates from air into organic solvents using the dynamic method was established using the solvent resistant optical sensor. The feasibility of online oxygen measurements in organic solvents has also been demonstrated, paving the way for new opportunities in process control.

Keywords: Optical sensor, organic solvents, oxygen transfer

Nomenclature: K_{La} , Oxygen mass transfer coefficient; $[O_2]^*$, Saturation concentration of oxygen; $[O_2]$, Concentration of oxygen at a given time; pO_2^* , Saturation partial

pressure of oxygen; **pO₂**, partial pressure of oxygen at a given time; **H**, Henry's Law constant; **DO**, dissolved oxygen

1 Introduction

Oxygen transfer is important in many bioprocesses employed at industrial scale. In aerobic fermentation and biocatalysis using growing cells, oxygen is used both for cell growth and maintenance. Likewise in biocatalysis, using resting cells or isolated enzymes for chemical synthesis, oxygen is used as a substrate in many reactions such as those catalysed by oxidases (1) or oxygenases (2–5). In all these biotechnological examples, the measurement of oxygen concentration in solution is essential, in order to provide valuable data which can subsequently be used for process monitoring and/or control (6).

Whilst most bioprocesses are carried out in an aqueous solution, there is increasing interest in the use of non-aqueous media to partially, or completely, replace water (7). This is motivated in large part by the significant benefits in process intensification that can result from the introduction of poorly water-soluble organic solvents above their water saturation concentration, resulting in a biphasic system (8–12). One particularly interestingly consequence of the introduction of an organic solvent is usually better oxygen transfer. Indeed it has even been suggested to use organic solvents as “oxygen vectors” in order to improve the oxygen supply in some fermentation processes (13,14). Oxygen transfer is improved due to the far higher solubility of oxygen in many organic solvents compared to water (15,16), where the solubility is limited to 5 mL O₂/L water (0% salinity, 30 °C) (17). Oxygen is usually supplied in air, meaning a maximum oxygen concentration in solution of about 250 μM oxygen. The higher oxygen solubility in organic solvents gives a higher concentration driving force for mass transfer, resulting in improved transfer rates. On this premise we reasoned that the

application of a sensor suitable for measurement of oxygen concentration in organic solvents would be useful.

Indeed, historically the determination of oxygen solubility in organic solvents has proven challenging. Methods reported include titration, use of manometers, photochemical methods, electron spin resonance and chromatography (18). However, none of these methods enable continuous measurement. For such an application, Clark-type electrochemical sensors have been established for the measurement of oxygen concentration in bioprocesses (19–21). However, despite their long history only a few studies report the use of Clark electrodes in apolar organic solvents (22) or alcoholic solutions. This is because the membrane used in a Clark electrode is typically made of a polymer which is prone to swelling when exposed to organic solvents, leading to erroneous measurements. More recently, the application of optical oxygen sensors has been reported as an alternative (23), using measurement principles based on luminescent quenching of an indicator dye by the molecular oxygen. The indicator dye is embedded in a host polymer composed of sol-gel, silicone and polystyrene. Nevertheless, most of the established optical sensors are not adequate for application in organic solvents either, due to the swelling or dissolution of the polymeric material. More recently some optical sensors with resistance to organic solvents were reported (24–28). On this basis we sought to use these sensors for continuous on-line measurement of oxygen concentration in organic solvents as the basis for the determination of oxygen transfer rates. This is we believe the first report of such measurements.

2 Experimental

All solvents, unless specified, were purchased in analytical grade from Sigma Aldrich (Steinheim, Germany) and used as purchased. The solvent-resistant oxygen sensors (model OXSOLV) and the FirestingO2 were obtained from PyroScience (Aachen, Germany).

2.1 Sensor characterization in organic solvents

Gas mixtures of defined oxygen partial pressure (pO_2) were obtained using two mass flow controllers (Read-Ysmart series) from Vögtlin instruments (www.voegtlin.com). Compressed air and pure nitrogen were used as calibration gases. The mixtures were controlled with an oxygen microsensor (model OXR230-CL4) from PyroScience. The calibration gas was passed through a glass trap and subsequently a glass measurement cell (maintained at 25° C in a water bath). In turn the glass trap and the measurement cell were filled with the organic solvents to be evaluated.

2.2 Reactor

Experiments were carried out in a 250 mL stirred tank reactor equipped with two Rushton turbines (6 blades, diameter 2.4 cm, width of the blade was 0.6 cm and height 0.5 cm) on a single drive shaft, and two baffles (of 1 cm width). Both the height of the impellers and the agitation speed of them were adjustable.

2.3 Calibration of the oxygen sensor

A 2-point calibration was made for the oxygen sensors by exposing them to air (for 100% DO) and 1% sodium sulphite (for 0% DO). The calibration for the zero reading was carried out with sodium sulphite since it was a fast and an easy-to-use calibration method.

2.4 K_La measurements for oxygen transfer from air to organic phase

k_La measurements were made for oxygen transferring from air to heptane, MTBE, isooctane and cyclohexane respectively, using the well-established dynamic gassing out

method (29). Nitrogen was used for gassing out and head-space aeration was employed as a means of oxygen supply.

50 mL of a given organic solvent was placed in the stirred tank reactor (maintained at 37°C in a water bath) and the stirring speed was set to 80 rpm for all solvents ensuring that the solvent was mixed without disturbing the surface. The low stirring speed ensured that there was no splashing and thus a fixed gas-liquid interfacial area was made available for oxygen transfer. The head-space was flushed with nitrogen until the oxygen concentration in the aqueous phase stabilized (zero reading). Subsequently, the nitrogen flow was stopped and the reactor left open to enable to fill the headspace. Oxygen then transferred from the headspace to a given organic solvent until the two were equilibrated. The increase in oxygen concentration in the respective organic solvent was measured using two oxygen sensors (one attached to each baffle) enabling duplicate measurements. For the calculation of the specific interfacial area (a) available for oxygen transfer, the volume of the solvents was considered to be 50 mL. Exceptionally in the case of MTBE, the volume used was 40 mL on account of evaporation losses during degassing, resulting in a higher specific interfacial area.

3 Results and discussion

In the following section, we report the measurement of oxygen concentration and transfer rate of oxygen from air to neat organic solvents. To that end, the oxygen sensor was first characterised in different solvents followed by use them in organic solvents with the aim of measuring oxygen transfer rates.

3.1 Characterisation of oxygen sensor in organic solvents

Initially, the oxygen sensor was tested in representative organic solvents, covering a wide range of different polarities. The active sensor coating is composed of a Teflon-like hydrophobic polymer, which makes it resistant to a range of non-aqueous solvents.

The sensor was exposed to the organic solvents for one hour. The sensors showed a response time (t_{90}) of 4 seconds, fast enough for the measurements required in the dynamic method. Figure 1 shows the calibration curves of the sensor at varying oxygen partial pressures, when exposed to the different organic solvents. The calibration curves are nearly identical for all tested solvents, confirming that the individual solvents do not influence the sensor material.

3.2 Measurement of oxygen transfer rate from gas to organic solvent

The sensors were then used to measure oxygen transfer rates from air to neat organic solvents. The selection of organic solvents was limited to those applicable in biological processes such as fermentation and biocatalysis. The solvents were selected therefore on the basis of a $\log P > 4$ (30) and which also met environmental guidelines, using the ACS GCI solvent selection guide (9).

Oxygen transfer rates from the headspace to the organic phase were measured by the well-established dynamic method (29). First, the oxygen concentration in the organic solvent was depleted by flushing the headspace with nitrogen (schematically illustrated in Figure 2). After flushing with nitrogen the open vessel filled with air, and the subsequent oxygen saturation profile was also obtained through headspace aeration. Bubbled aeration was avoided since it would lead to solvent evaporation, complicating interpretation of the results.

When the nitrogen sweeps out the air from the headspace, the oxygen concentration becomes zero. When headspace aeration follows, the oxygen concentration rises in the system until it reaches the saturation concentration (as seen in Figure 3). It is important that the sensors used for k_{La} measurements have a fast response in order to avoid any errors due to artefacts as a result of the measurement itself (31).

The oxygen transfer rate can be described by equation 1:

$$\frac{d[O_2]}{dt} = k_L a * ([O_2]^* - [O_2]) - r_{uptake} \quad (1)$$

Where $d[O_2]/dt$ represents the rate of change of oxygen concentration, t time, $[O_2]^*$ and $[O_2]$, the saturation concentration of oxygen and the concentration of oxygen in the liquid phase at a given time, respectively. r_{uptake} represents the rate of oxygen uptake in the system. Since there is no active consumption of oxygen, the oxygen uptake rate is zero and the equation can be reduced to equation 2:

$$\frac{d[O_2]}{dt} = k_L a * ([O_2]^* - [O_2]) \quad (2)$$

The volumetric mass transfer coefficient is represented by “ $k_L a$ ” which is itself composed of “ k_L ” representing the mass transfer coefficient and “ a ” the specific interfacial (32).

According to Henry’s law, the concentration of oxygen in solution in equilibrium with the gas, $[O_2]^*$, is proportional to the partial pressure of oxygen in the gas (pO_2^*)(equation 3):

$$[O_2]^* = H * pO_2^* \quad (3)$$

Re-writing equation 2 in terms of partial pressure (pO_2) and saturation partial pressure (pO_2^*) gives equation 4:

$$\frac{HpO_2}{dt} = k_L a * (HpO_2^* - HpO_2) \quad (4)$$

Solving equation 4 gives equation 5:

$$\ln(pO_2^* - pO_2) = -k_L a * t + \ln pO_2^* \quad (5)$$

In this way, the $k_L a$ for each solvent can be obtained from the plot of $\ln(pO_2^* - pO_2)$ as a function of time. Partial pressures were measured using the sensors (Figure 3), pO_2^* is the partial pressure recorded at equilibrium and pO_2 , the partial pressure recorded at a given time.

k_L values were calculated from the estimated $k_L a$ for the different solvents and are presented in Table 1. As a control water was also tested and from Table 1, it can be seen that the mole fraction of oxygen in the water phase is lower by an order of magnitude than in any of the organic solvents reflecting the higher oxygen solubility measured in the organic solvents (16). As expected solubility of oxygen in the solvents reflects the P_{sat}^* values. The results indicate not only that the sensor can be used to measure oxygen in organic solvent, but can also be used on-line in a dynamic method, thus facilitating the collection of oxygen transfer data. To the best of our knowledge this is the first time such results are reported. Such online measurements can be particularly useful when operating a biological system in a Lewis cell (33–35) to measure the mass transfer that occurs between two immiscible liquid phases with a defined interfacial area. The k_L values obtained from the Lewis cell experiments can be used to predict the oxygen transfer rates for a given phase ratio for a well-mixed large scale system. For example, if an oxygen transfer rate from organic solvent to aqueous phase (where the oxygen requiring biocatalytic process occurs) is required, the solvent sensors could be placed in each of the phases (aqueous and organic) and oxygen concentrations measured over time. The result obtained would in turn indicate the mass transfer rates from gas to organic phase and subsequently organic phase to aqueous phase.

4 Concluding remarks

Oxygen transfer rates from air to neat organic solvents were determined with a solvent resistant optical sensor. The study demonstrates for the first time the feasibility of online measurements of oxygen in organic solvents, which was previously limited by unstable sensors. The sensor therefore presents new possibilities in process control of

intensified for fermentation and biocatalytic processes where changes in oxygen concentrations occur in the presence of organic solvents.

Acknowledgement

Financial support by the European Union FP7 Project BIOINTENSE - Mastering Bioprocess integration and intensification across scales (Grant Agreement Number 312148) is gratefully acknowledged.

Conflicts of interest

The authors declare that there are no conflicts of interest.

5 References

1. Li T, Liang J, Ambrogelly A, Brennan T, Gloor G, Huisman G, et al. Efficient, chemoenzymatic process for manufacture of the boceprevir bicyclic [3.1.0]proline intermediate based on amine oxidase-catalyzed desymmetrization. *J Am Chem Soc.* 2012; 134:6467–72.
2. Huisman GW, Collier SJ. On the development of new biocatalytic processes for practical pharmaceutical synthesis. *Curr Opin Chem Biol.* 2013; 17:284–92.
3. Bong Y, Michael C, Stevem C, Benjamin M, Vogel M, Zhang X, et al. Synthesis of prazole compounds. USA: Codexis Inc.; WO2011071982, 2011.
4. Baldwin CVF, Wohlgemuth R, Woodley JM. The first 200-L scale asymmetric Baeyer-Villiger oxidation using a whole-cell biocatalyst. *Org Process Res Dev.* 2008; 12:660–5.
5. Furuya T, Miura M, Kuroiwa M, Kino K. High-yield production of vanillin from ferulic acid by a coenzyme-independent decarboxylase/oxygenase two-stage process. *N Biotechnol.* Elsevier B.V.; 2015; 32:335–9.
6. Demarche P, Junghanns C, Ardao I, Agathos SN. Dynamic measurement of oxidase activity based on oxygen consumption in open systems. *Eng Life Sci.* 2015; 1–11.
7. Marques MPC, De Carvalho CCCR, Cabral JMS, Fernandes P. Scaling-up of complex whole-cell bioconversions in conventional and non-conventional media. *Biotechnol Bioeng.* 2010; 106:619–26.
8. Liang J, Lalonde J, Borup B, Mitchell V, Mundorff E, Trinh N, et al. Development of a biocatalytic process as an alternative to the (-)-DIP-Cl-mediated asymmetric

- reduction of a key intermediate of montelukast. *Org Process Res Dev.* 2010; 14:193–8.
9. Reslow M, Adlercreutz P, Mattiasson B. Organic solvents for bioorganic synthesis. *Appl Microbiol Biotechnol.* 1987; 26:1–8.
 10. Kuhn D, Kholiq MA, Heinzle E, Bühler B, Schmid A. Intensification and economic and ecological assessment of a biocatalytic oxyfunctionalization process. *Green Chem.* 2010;12 :815.
 11. Bruno B, Schmid A. Process implementation aspects for biocatalytic hydrocarbon oxyfunctionalization. 2004; 113:183–210.
 12. Schmölzer K, Mädje K, Nidetzky B, Kratzer R. Bioprocess design guided by in situ substrate supply and product removal: Process intensification for synthesis of (S)-1-(2-chlorophenyl)ethanol. *Bioresour Technol.* 2012; 108:216–23.
 13. Dumont E, Andrès Y, Le Cloirec P. Effect of organic solvents on oxygen mass transfer in multiphase systems: Application to bioreactors in environmental protection. *Biochem Eng J.* 2006; 30:245–52.
 14. Nielsen DR, Daugulis AJ, McLellan PJ. A restructured framework for modeling oxygen transfer in two - phase partitioning bioreactors. *Biotechnol Bioeng.* 2005; 91:773–7.
 15. Fischer K, Wilken M. Experimental determination of oxygen and nitrogen solubility in organic solvents up to 10 MPa at temperatures between 298 K and 398 K. *J Chem Thermo.* 2001; 33 :1285–308.

16. Battino R, Rettich TR, Tominaga T. The solubility of oxygen and ozone in liquids. *J Phys Chem Ref data*. 1983; 12 :163–78.
17. Murray CN, Riley JP. The solubility of gases in distilled water and sea water - II. Oxygen. *Deep Res*. 1969; 16:311–20.
18. Quaranta M, Murkovic M, Klimant I. A new method to measure oxygen solubility in organic solvents through optical oxygen sensing. *Analyst*. 2013; 138: 6243–5.
19. Clark LC. Electrochemical device for chemical analysis. US Patent 2913386. 1959.
20. Chang SC, Stetter JR, Cha CS. Amperometric gas sensors. *Talanta*. 1993; 40:461–77.
21. Suresh S, Srivastava VC, Mishra IM. Techniques for oxygen transfer measurement in bioreactors: a review. *J Chem Technol Biotechnol*. 2009; 84:1091–103.
22. Escobar L, Salvador C, Contreras M, Escamilla JE. On the application of the Clark oxygen electrode to the study of enzyme kinetics in apolar solvents: the catalase reaction. *Anal Biochem*. 1990; 184:139–44.
23. Wang X, Wolfbeis OS. Optical methods for sensing and imaging oxygen: materials, spectroscopies and applications. *Chem Soc Rev*. 2014; 43:3666–761.
24. Costa-Fernández JM, Diaz-García ME, Sanz-Medel a. Sol-gel immobilized room-temperature phosphorescent metal-chelate as luminescent oxygen sensing material. *Anal Chim Acta*. 1998; 360:17–26.
25. González-Carrero S, de la Guardia M, Galian RE, Pérez-Prieto J. Pyrene-Capped CdSe@ZnS Nanoparticles as Sensitive Flexible Oxygen Sensors in Non-Aqueous Media. *Chemistry Open* 2014; 3:199–205.

26. García-Fresnadillo D, Marazuela MD, Moreno-Bondi MC, Orellana G. Luminescent Nafion membranes dyed with ruthenium(II) complexes as sensing materials for dissolved oxygen. *Langmuir*. 1999; 15:6451–9.
27. Velasco-García N, Valencia-González MJ, Díaz-García ME. Fluorescent organofilms for oxygen sensing in organic solvents using a fiber optic system. *Analyst*. 1997; 122:1405–9.
28. Xavier P, Garci D. Oxygen Sensing in Nonaqueous Media Using Porous Glass with Covalently Bound Luminescent Ru (II) Complexes. 1998; 70:5184–9.
29. Garcia-Ochoa F, Gomez E. Bioreactor scale-up and oxygen transfer rate in microbial processes: an overview. *Biotechnol Adv*. 2009; 27:153.
30. León R, Fernandes P, Pinheiro HM, Cabral JMS. Whole-cell biocatalysis in organic media. *Enzyme Microb Technol*. 1998; 23:483–500.
31. Tribe LA, Briens CL, Margaritis A. Determination of the volumetric mass transfer coefficient (kLa) using the dynamic “gas out–gas in” method: Analysis of errors caused by dissolved oxygen probes. *Biotechnol Bioeng*. 1995; 46:388–92.
32. Villadsen J, Nielsen J, Lidén G. *Bioreaction engineering principles*. Third Edition, Springer; 2011.
33. Bauer M, Geyer R, Griengl H, Steiner W. The use of Lewis cell to investigate the enzyme kinetics of an (S)-hydroxynitrile lyase in two-phase systems. *Food Technol Biotechnol*. 2002; 40:9–19.

34. Lewis JB. The mechanism of mass transfer of solutes across liquid-liquid interfaces: Part I: the determination of individual transfer coefficients for binary systems. Chem Eng Sci. 1954; 3:248–59.
35. Woodley JM, Brazier AJ, Lilly MD. Lewis cell studies to determine reactor design data for two - liquid - phase bacterial and enzymic reactions. Biotechnol Bioeng. 1991; 37:133–40.

Table 1. $K_L a$ and K_L values for different solvent at 30 °C.

Solvent	$k_L a$ (h⁻¹)	k_L (mh⁻¹)	P^*_{sat} (mbar)	Mol fraction of oxygen*1000^a
MTBE	1.73	0.021 ^b	150	-
Cyclohexane	2.56	0.039	160	1.24
Heptane	3.96	0.060	185	2.11
Isooctane	2.88	0.043	189	2.63
Water	7.2	0.108	220	0.02

^aAs reported by 16

^bThe specific surface interfacial area used for MTBE is calculated with the volume of the liquid to be 40 mL (the reduced volume considered here is to account for evaporation of solvent during the deoxygenation phase).

Figure legends




Figure 1. Calibration curves of oxygen sensors in different organic solvents at 25 °C.]





Figure 2. Photograph of the experimental set up. The dimensions of the rotors and baffles are represented on the left and the assembled reactor on the right.






Figure 3. (a) Dynamic method for determination of oxygen transfer rates from the headspace to heptane system. (b) Dynamic method for oxygen transfer measurements from isooctane, cyclohexane, MTBE and heptane.